

**OAR 333-030-0105**

8) AUTOMATIC EXTERNAL DEFIBRILLATOR (AED). By June 1, 2009, all residential camps with 100 or more campers and staff on-site at any one time must have at least one AED with pediatric capability, located at the camp. The camp operator must comply with the following:

- (a) Each AED must have documented maintenance inspections and service records, including the battery and electrodes according to the guidelines set forth by the manufacturer.
- (b) Equipment to be stored with the AED must include scissors, CPR face mask, protective gloves, a disposable razor for removing excess hair, and an absorbent cloth or pad.
- (c) There must be a sufficient number of staff trained in the use of the AED so that there is at least one trained adult on-site whenever the camp is used for more than three consecutive nights for the normal camp programs or by a contract or rental group.
- (d) The AED must be stored in a central location where the AED is accessible and can be quickly retrieved.
- (e) Signage must be provided that indicates the location of the AED.
- (f) A policy must be developed for the use of the AED, including the need to contact 911 as soon as possible. This policy should be made available to camp staff and must be posted with the AED.

**Note:** It is recommended that all camps consider obtaining AEDs, in compliance with sections (5)(a) through (5)(f), as their budget allows.